

Independent

INSTRUCTIONS FOR USE

Codes 1545-1547




CONTENTS	PAGE
1.0 INTRODUCTION	1
2.0 ILLUSTRATION OF YOUR INDEPENDENT STANDER	2
3.0 FOR YOUR SAFETY	3
4.0 UNPACKING YOUR INDEPENDENT STANDER	3
5.0 SETTING UP AND ADJUSTING YOUR INDEPENDENT STANDER	4
6.0 CARE & MAINTENANCE	7
7.0 GUARANTEE & SERVICE	8
8.0 CONTINUOUS IMPROVEMENT	9
9.0 WARRANTY & AFTERSALE	10

1.0 INTRODUCTION


Thank you for choosing the Smirthwaite Independent vertical stander.

The Independent is our highly adjustable vertical stander designed for children from 18 months old to young adults who require moderate levels of support when weight bearing. Like the Wentworth, its design is ideal for home and multi-user environments such as nurseries and schools.


	<p>IMPORTANT!</p> <p>These instructions should be read by all therapists and carers using the equipment and should be retained for future reference. The product should always be used under adult supervision.</p> <p>Any incorrect use of the product and failure to follow the instructions may put the user at risk or impede the function. If you have any queries using this product or wish for further copies, please do not hesitate to contact Customer Service department on T: +44 (0) 1626 835552.</p>
---	--

2.0 ILLUSTRATION OF YOUR INDEPENDENT STANDER

- A. Castor
- B. Tray
- C. Knee blocks
- D. Pelvic harness
- E. Lateral supports
- F. Footplate & Sandals



3.0 FOR YOUR SAFETY

	STOP!
	Please read these instructions CAREFULLY and THOROUGHLY.


- The user should NOT be left unattended whilst in the stander. Always ensure a responsible therapist or carer is in attendance.
- If you believe the stander or any fitted accessory to be faulty at any time, **DO NOT USE** – contact Smirthwaite by telephone on +44 (0) 1626 835552.
- The therapist should be familiar with the methods of adjustment and have completed all adjustments appropriately to meet the needs of the child before transferring the child into the stander.
- The stander is **ONLY** to be used indoors on a flat level surface.
- Always fasten the straps and harnesses provided with the stander.
- Regular maintenance checks and cleaning are essential for the safe use of this equipment (see Section 6.0 Care and Maintenance).
- Always keep this product away from naked flames, cigarettes and sources of heat including open fireplaces, radiators and heaters.
- **DO NOT** fit parts or accessories of other manufacturers to this product unless authorised to do so in writing by G&S Smirthwaite Ltd. Failure to follow these instructions will not only invalidate the guarantee but could make the stander dangerous to use. G&S Smirthwaite Ltd will not accept liability for any injury or damage incurred through such malpractices. Any repairs required must be carried out by G&S Smirthwaite Ltd authorised personnel.

4.0 UNPACKING YOUR INDEPENDENT STANDER

- Your stander will be delivered assembled and will simply require final adjustment to suit the child's clinical needs before use.
- Once the stander has been carefully unpacked, please check all parts.

If you believe this product to be faulty - **DO NOT USE** - Contact G&S Smirthwaite Ltd on T: +44 (0) 1626 835552.



5.0 SETTING UP AND ADJUSTING YOUR INDEPENDENT STANDER

	STOP!
	<p>If in any doubt, ALWAYS seek ADVICE.</p> <p>Always turn handwheels, levers and screws clockwise to tighten or anti-clockwise to loosen. The stander must be fully adjusted by a therapist or trained representative before use.</p>

5.1 CHASSIS EXTENSION

To ensure stability, the chassis has castor extensions, located in the box end sections (A).


- Loosen the tightening screw located in the box section (B), using the hexagonal key provided.
- Set the extension to the desired length.
- Re-tighten screw (B).
- Repeat for the three remaining extensions.
- Check stability of the stander before allowing a child to use the product.



5.2 OVERALL HEIGHT

The overall height of the stander can be adjusted to set the tray height.

- Loosen the lever locks (2 off) on the angled support legs (C).
- Loosen the lever locks (2 off) on the main support legs (D).
- Lift the main support legs to the desired height and then tighten (D).
- Set the tray level by lifting the angled support legs, and tighten (C).



5.3 TRAY ANGLE

The tray can angle adjusted as if required.

- Loosen the handwheels (2 off) on each side of the tray (E).
- Two fixing holes are available to angle the tray – select the most appropriate for the child’s needs.
- Re-tighten the handwheels (2 off) on each side of the tray (E).



5.4 KNEE BLOCK

The knee blocks can be adjusted for height, depth, width and angle.


5.4.1 HEIGHT

- Loosen the lever lock (F) on each main support leg.
- Slide the knee block assembly to the desired height.
- Re-tighten the lever locks (F).



5.4.2 DEPTH AND ANGLE

- Loosen the handwheels (G) on each side of the knee block assembly.
- The assembly depth can now be adjusted backward or forward.
- By twisting the assembly, the angle can be adjusted.
- When the desired depth/angle has been set, re-tighten the handwheels (G).



5.4.3 WIDTH


- To adjust the width (pitch) of the knee cups, simply loosen the handwheel at the rear of each knee block (H) and slide to the desired position.
- The amount of adjustment is limited by the slot.
- Re-tighten each handwheel when finished.

5.5 PELVIC HARNESS

The pelvic harness can be adjusted for height, depth and angle.

5.5.1 HEIGHT

- Loosen the lever lock (I) on each main support leg.
- Slide the pelvic harness assembly to the desired height.
- Re-tighten the lever locks (I).



5.5.2 DEPTH AND ANGLE

- Loosen the handwheels (J) on each side of the pelvic harness assembly.
- The assembly depth can now be adjusted backward or forward.
- By twisting the assembly, the angle can be adjusted.
- When the desired depth/angle has been set, re-tighten the handwheels (J).

5.6 LATERAL SUPPORT

The lateral support can be adjusted for height, depth and angle.

5.6.1 HEIGHT

- Loosen the lever lock (K) on each main support leg.
- Slide the lateral support assembly to the desired height.
- Re-tighten the lever locks (K).



5.6.2 DEPTH AND ANGLE

- Loosen the handwheels (L) on each side of the lateral support assembly.
- The assembly depth can now be adjusted backward or forward.
- By twisting the assembly, the angle can be adjusted.
- When the desired depth/angle has been set, re-tighten the handwheels (L).

5.6.3 WIDTH

- To adjust the width (pitch) of the lateral support, simply loosen the handwheels at the rear of the lateral support and slide to the desired position.
- The amount of adjustment is limited by the slot.
- Re-tighten each handwheel when finished.


5.7 ACTIVITY BOWL

- When required the bowl fits in the tray. When not in use, the tray infill can be used.

5.8 SANDALS

- The child's feet can be strapped whilst sitting prior to transfer from a chair.
- The sandals can be adjusted in the horizontal plane.
- Vertical and angular alterations can be made by using step blocks under the sandals – please contact the factory or your local product advisor for further details.


6.0 CARE AND MAINTENANCE

	IMPORTANT!
	Cleaning is recommended on a regular basis

- Clean upholstery, metal and wooden parts with a damp cloth and mild detergent.
- Stubborn marks on wood work should be cleaned by using a soft brush.
- Do not soak or immerse the stander in water.
- Store the stander in a cool dry place and out of direct sunlight.
- DO NOT use bleach, solvents, abrasives, synthetic detergents, wax polishes, antibacterial sprays or wipes.
- For further information please refer to MHRA or your NHS cleaning guidelines.

6.1 DAILY CHECKS

- Check all parts for signs of wear and tear or damage.
- Check all straps for fraying, and that buckles are not missing/damaged.
- Check all screws, lever locks, and handwheels are present and lock satisfactorily.

	STOP!
	The user should not be seated while the checks are carried out.

6.2 SERVICE INTERVAL

The Independent vertical stander should be serviced every 3 years. Servicing must only be undertaken by a Smirthwaite service engineer, or by a Smirthwaite trained representative.

7.0 GUARANTEE & SERVICE

The stander is issued with a full parts and labour guarantee for 2 years from the date of delivery. This guarantee does not apply to accidental damage caused through inappropriate use of the stander. This guarantee is issued at the discretion of G&S Smirthwaite Ltd on RTB (Return to Base) basis.

If you have any questions or doubts relating to the safety or use of the stander please contact us for advice or assistance on:

T: +44 (0) 1626 835552

E: info@smirthwaite.co.uk

8.0 CONTINUOUS IMPROVEMENT

G&S Smirthwaite Ltd are committed to continuous improvement to their product range. Should you have any suggestions or comments please send them to our product design department at: info@smirthwaite.co.uk

G&S Smirthwaite Ltd reserve the right to change the specification or material without prior notice.

For catalogues, help and further information on our products please contact us at:

G&S Smirthwaite Ltd
17 Wentworth Road
Heathfield
Devon TQ12 6TL

T: +44 (0) 1626 835552

F: +44 (0) 1626 835428

E: info@smirthwaite.co.uk

W: www.smirthwaite.co.uk



G&S Smirthwaite, 17 Wentworth Road, Heathfield, Newton Abbot, Devon.TQ12 6TL
T: +44 (0) 1626 835552 F: +44 (0) 1626 835428 E: info@smirthwaite.co.uk
W: www.smirthwaite.co.uk